

High-precision measuring instrument for determining the concentration of SF₆ gas

Model GA65

技术咨询和询价：010-68940148

WIKA data sheet SP 62.13

康高特-WIKA GA65 SF6气体综合测试仪

SF₆ Tracer

Applications

- Leak test for the final inspection of SF₆ gas-filled equipment
- Monitoring of the concentration of SF₆ gas in the ambient air of enclosed spaces

Special features

- High-precision and reproducible measurements in the ppb range
- Fast response time
- Simple operation and long service intervals
- No consumables, e.g. flush gas, are required
- Expendable by multiplexer for up to 24 measuring points



Leak rate measuring instrument for SF₆ gas, model GA65

Description

The model GA65 measuring instrument has been specifically designed for measuring small concentrations of SF₆ gas. The quantitative measurement of SF₆ gas in the air is carried out reliably and reproducibly even at the smallest quantities.

The used technology is based on the photo-acoustic infrared spectroscopy. This physical and non-destructive measuring principle achieves a very high accuracy with a detection rate of 6 ppb_v.

Humidity is compensated and thus does not influence the measuring result.

Cyclic self tests guarantee the reliability and functionality of the instrument. It is recommended to recalibrate the instrument once a year.

The leak rate measuring instrument is easy to use and can be operated via control keys at the front of the housing or via an extensive PC software with a graphic user interface.

Both operating modes allow for the setting of the parameters (e.g. duration of the sampling), the starting of a measurement (manually or automatically), the display of the concentration of SF₆ gas in real time or the sending of the values to the downstream control software.

Specifications

Measuring principle

Photo-acoustic infrared spectroscopy

Detection limit

6 ppb_v or 6 x 10⁻⁹ ml/s
(at a flow rate of 60 ml/min)

Measuring range

6 ... 60,000 ppb_v

Resolution

1 ppb_v

Sensor characteristics

Temperature and pressure compensated
Humidity: Cross compensated up to 80 % and 31 °C

Reproducibility

1 %

Response time t₉₀

approx. 15 seconds

Permissible temperature ranges

Operation: 5 ... 40 °C
Storage: -25 ... +55 °C

Service interval

Once a year

Warning signals

2 settable alarm values
Audible and visible

Electrical output

2 relays (settable alarm values)

Data storage

Available (internal storage space)
Software and connection cable included in delivery

Voltage supply

AC 100 ... 240 V, 45 ... 67 Hz, 70 W

Interface

IEEE-488 and RS-232

Dimensions

W x H x D: 395 x 175 x 300 mm

Weight

9 kg

Ingress protection

IP 20

Ordering information

The description of the model is sufficient.

© 2013 WIKA Alexander Wiegand SE & Co. KG, all rights reserved.
The specifications given in this document represent the state of engineering at the time of publishing.
We reserve the right to make modifications to the specifications and materials.

Drucksachenänderung/Modification notice

Dokument/Document Data sheet SP 62.13
Sprache/Language GB
Ausgabe/Version 04/2013
Dateiname/File name DS_SP6213_GB

Beiliegend erhalten Sie die aktuelle Ausgabe der oben genannten Technischen Dokumentation, die vorherige Ausgabe wird hiermit ungültig./Enclosed please find the current version of the above mentioned document. The previous edition becomes invalid.

Folgende Änderungen wurden durchgeführt/Following modifications have been made:

Seite Page	Änderungen (Text, Fotos, Zeichnungen) Modifications (text, photo, drawings)
1-2	Completely reworked

Ablagehinweis für den WIKAI Gesamtkatalog, Rubrik
Sorting information for WIKAI Full Catalogue, section

Bitte teilen Sie Ihren Mitarbeitern diese Änderungen mit.
Please forward these modifications to your colleagues.

Bearbeitet/Modified			Geprüft/Checked			Freigegeben/Approved		
5.4.2013	MS	E. Lungavita	5.4.2013	SF6	O.Rieger	5.4.2013	MS	N. Kroth

History data sheet SP 62.15 GB

Sprache: GB
Seitenzahl: 3
Erstellt von: E. Lungavita
Abt.: MS
am: 5.4.2013
Ausgabe: 04/2013

Version	Pages	Changes	Person	Date	Release date
a		see folder (in Archive)			07/2010
b		see folder (in Archive)			07/2010
c	1-3	Completely reworked	E. Lungavita	5.4.2013	04/2013